

University of Stuttgart
Germany



Analytische Qualitätssicherung Baden-Württemberg

**Proficiency Test 5/16
TW S2 – pharmaceuticals
in drinking water**

**sulfadiazine, sulfadimidine, sulfamethoxazole,
sulfaethoxypyridazine, sulfamerazine, sulfathiazole
sulfadoxine, sulfamethoxypyridazine,
sulfachloropyridazine, sulfadimethoxine,
trimethoprim**

provided by
AQS Baden-Württemberg at
Institute for Sanitary Engineering, Water Quality and Solid Waste Management,
University of Stuttgart
Bandtäle 2, 70569 Stuttgart-Büsnau, Germany

and
IWW Water Center
Moritzstr. 26, 45476 Mülheim an der Ruhr, Germany



Stuttgart, in December 2016

Responsibilities:

Scientific director AQS: Dr.-Ing. Dipl.-Chem. Michael Koch

PT manager: Dr.-Ing. Frank Baumeister

AQS Baden-Württemberg at
Institute of Sanitary Engineering,
Water Quality and Solid Waste Management
at University of Stuttgart

Bandtäle 2

70569 Stuttgart-Büsnau

Germany

<http://www.aqsbw.de>

Tel.: +49 (0)711 / 685-65446

Fax: +49 (0)711 / 685-63769

E-Mail: info@aqsbw.de

LIST OF CONTENTS

GENERAL	1
PT DESIGN	1
SAMPLE PREPARATION	1
SAMPLE DISTRIBUTION	1
ANALYTICAL METHODS	2
SUBMISSION OF RESULTS	2
EVALUATION PROCEDURE	2
ASSESSMENT	3
EVALUATION	3
EXPLANATION OF APPENDIX A	4
EXPLANATION OF APPENDIX B	5
EXPLANATION OF APPENDIX C	5
MEASUREMENT UNCERTAINTY	6
TRACEABLE REFERENCE VALUES	7
INTERNET	8

Appendix A

SULFADIAZINE	A-1
SULFAMIDINE	A-7
SULFAMETHOXAZOLE	A-13
SULFAETHOXYPYRIDAZINE	A-19
SULFAMERAZINE	A-25
SULFATHIAZOLE	A-31
SULFADOXINE	A-37
SULFAMETHOXYPYRIDAZINE	A-43
SULFACHLOROXYPYRIDAZINE	A-49
SULFADIMETHOXINE	A-55
TRIMETHOPRIME	A-61

Appendix B**Appendix C**

SULFADIAZINE	C-1
SULFAMIDINE	C-10
SULFAMETHOXAZOLE	C-19
SULFAETHOXYPYRIDAZINE	C-28
SULFAMERAZINE	C-37
SULFATHIAZOLE	C-46
SULFADOXINE	C-55
SULFAMETHOXYPYRIDAZINE	C-64
SULFACHLOROXYPYRIDAZINE	C-73
SULFADIMETHOXINE	C-82
TRIMETHOPRIME	C-91

General

This PT was provided by AQS Baden-Württemberg in cooperation with IWW Water Center in Mülheim an der Ruhr and with the network "NORMAN" (Network of reference laboratories for monitoring of emerging environmental pollutants). In this round sulfadiazine, sulfadimidine, sulfamethoxazole, sulfaethoxypyridazine, sulfamerazine, sulfathiazole, sulfadoxine, sulfamethoxypyridazine, sulfachloropyridazine, sulfadimethoxine and trimethoprim were to be determined.

The PT was executed and evaluated according to the requirements of DIN 38402-A45 and ISO/TS 20612.

PT design

Each participant received the following samples:

- 3 samples for the determination of above mentioned parameters in 1000-ml-ground bottles. The samples were preserved by adding 40 mg/l sodium azide. The samples also contained acetonitrile as solubility promoter.

3 different concentration levels/batches were produced. All participants received the same samples.

Sample preparation

The samples for the determination of the pharmaceuticals were based on a real ground water matrix from the northern part of the region Ruhr in North Rhine-Westphalia. The ground water was used without treatment for the sample preparation.

The ground water was spiked with stock solutions and the concentrations covered drinking and ground water relevant ranges.

Sample distribution

The samples were dispatched on 11 October 2016 by express service.

Analytical methods

The participants were free to choose a suitable method, but following limits of quantification were required:

Parameter	limit of quantification
• sulfadiazine	0.05 µg/l
• sulfadimidine	0.05 µg/l
• sulfamethoxazole	0.05 µg/l
• sulfaethoxypyridazine	0.05 µg/l
• sulfamerazine	0.05 µg/l
• sulfathiazole	0.05 µg/l
• sulfadoxine	0.05 µg/l
• sulfamethoxy-pyridazine	0.05 µg/l
• sulfachloropyridazine	0.05 µg/l
• sulfadimethoxine	0.05 µg/l
• trimethoprim	0.05 µg/l

The samples had to be analysed in duplicate over the complete method (sample preparation and measurement). The participants were asked to report the results as average values in µg/l with three significant digits.

Submission of results

The deadline for the submission of results was on 07 November 2016.

Evaluation procedure

The statistical evaluation was executed according to DIN 38402-A45 and ISO TS 20612 "Interlaboratory comparisons for proficiency testing of analytical chemistry laboratories". From the participants' results a relative standard deviation was calculated for each concentration level and parameter using the Q-method. The reference values (see chapter "Traceable reference values") were used as assigned values x_{pt} . The standard deviation resulting from the Q-method was used as σ_{pt} .

σ_{pt} was limited as follows:

- lower limit: 5%
- upper limit: 25%

A z-score was calculated for each measurement result derived from the assigned value x_{pt} and the standard deviation for proficiency assessment σ_{pt} :

$$z\text{-Score} = \frac{(x - x_{pt})}{\rho_{pt}}$$

The z-score was modified to a z_U -score with a correction factor for proficiency assessment (as described in the standards mentioned above).

The tolerance limit was defined as $|z_U|=2.0$.

Assessment

There was no overall assessment of the proficiency test round, but the single parameters were assessed.

A parameter was assessed as successful, if more than half of the values were correctly determined (2 out of 3 values are within the tolerance limits).

According to ISO 13528 (2015) the single results were assessed as follows:

$ z_u \leq 2.0$	successful
$2.0 < z_u < 3.0$	questionable
$ z_u \geq 3.0$	unsatisfactory

Not successful were:

- 1) Values which were not determined (if the other samples of this parameters were analysed),
- 2) Values, which were indicated with "lower than limit of quantification",
- 3) Values, which were subcontracted,
- 4) Values, which were submitted after the deadline of submission of results.

Evaluation


Number of participants: 26

1 laboratory did not report any result.

Number of reported values: 621

Number of accepted values: 544 (87,6%)

In the following figure the successful and not successful laboratories for each parameter are illustrated.



Explanation of Appendix A

Appendix A contains for each parameter

- parameter tables
- a figure of participants' means versus the spiked amounts for the determination of the recovery rate
- a figure of the relative standard deviations versus the concentrations
- a figure of the tolerance limits in the PT versus the concentrations
- the frequency of application of analytical methods
- the method specific evaluation
- a comparison of mean and reference values for each concentration level
- a comparison of the relative standard deviations of the different methods
- the statistical characteristics of the method specific evaluation
- a tabular comparison of the means with the reference values and their uncertainties

Parameter tables

In these tables the following values for each concentration level are listed:

- assigned value
- expanded uncertainty of the assigned value in %, calculated from an uncertainty budget (see chapter "Traceable reference values")
- standard deviation, calculated using robust statistical method
- standard deviation for proficiency assessment for the calculation of z_U -scores
- rel. standard deviation for proficiency assessment
- tolerance limits above and below
- permitted deviations above and below in %
- number of values in this level
- number of not satisfactory values below and above the assigned value and the percentage of these values in total.

Determination of recovery rate

In the diagrams of the assigned values versus the spiked amount of analyte a linear regression line was calculated using a generalized least square regression which takes into account the uncertainties of the values in both directions. From these values the recovery rate for each parameter was determined (see diagrams). The slope of the line indicates the average recovery rate. The diagrams also contain the expanded uncertainty ($k=2$) of the concentrations from the spike and the assigned values.

Relative standard deviations and tolerance limits

The diagrams for the relative standard deviation vs. the assigned value show the concentration dependency of the standard deviation and the tolerance limits in percent. The relative standard deviations calculated from participants' data are the stars connected by an interrupted line, the rel. standard deviation taken from the variance function (and sometimes limited by the upper or lower limit) are given by squares, connected by a continuous line.

Method specific evaluation

For each parameter the methods used by the participants are shown in a diagram. In a second diagram for each method with a frequency of more than 5 %, values are sorted in 5 categories:

too low	results with z_U -score < -2
low	results with $-2 \leq z_U$ -score < -1
correct	results with $-1 \leq z_U$ -score $\leq +1$
high	results with $+1 < z_U$ -score $\leq +2$
too high	results with z_U -score $> +2$

Comparison of means and reference values for each concentration level

Finally the mean value calculated from all results, the reference value (see chapter Traceable reference values) are compared with mean values calculated for all methods separately. All mean values were calculated using the Hampel estimator described in ISO/TS 20612. Mean values were calculated only, if more than 8 results were within a z-score-range of ± 2 . The means are reported with their expanded uncertainty, calculated according to ISO 13528.

All mean values and their expanded uncertainties are additionally compared with the reference values and their expanded uncertainties.

Explanation of Appendix B

Participants were asked to report expanded uncertainties of their results on a voluntary basis. In this diagram for each parameter the reported uncertainties for all concentration levels with the reproducibility standard deviation (horizontal line) are displayed. Values which deviate from the reproducibility standard deviation with a factor more than 2 are usually not realistic.

Explanation of Appendix C

In the last part of the report, for all concentration levels the results of all participants are illustrated. Confidentiality of participants is ensured by using lab codes. The lab codes were sent to participants with the certificates.

In detail Appendix C contains:

- a table with all data
- figures with
 - o all reported results
 - o all z_U -scores
 - o all reported expanded uncertainties
 - o all ζ -scores

Table with all data

The assigned value with the expanded uncertainty and the tolerance limits for the concentration level is illustrated in the table. For each participant the following data are given:

- lab code
- reported result
- measurement uncertainty of the value (if reported)
- ζ -score for this value, calculated with the following formula

$$\zeta = \frac{x - x_{pt}}{\sqrt{u_{lab}^2 + u_{x_{pt}}^2}}, \text{ with}$$

$x - x_{pt}$ = difference from the measured value and the assigned value

- u_{lab} = standard uncertainty of the value, reported by the participant
- $u_{x_{pt}}$ = standard uncertainty of the assigned value

- z_U -score for proficiency assessment
- assessment of the value according to its z-score

Meaning of ζ -scores:

The assessment of ζ -scores is similar to that of z_U -scores. If the data are normally distributed and the uncertainties are well estimated, ζ -scores will lie between -2 and +2 with a probability of around 95 %.

ζ -scores are mainly influenced by the measurement uncertainties reported by the laboratory. Therefore ζ -scores are usually not appropriate for the assessment of the reported results, unless the reported measurement uncertainty is checked for fitness-for-purpose.

Therefore we do not use the ζ -scores for the assessment of the laboratories.

Nevertheless ζ -scores are appropriate to check the plausibility of the reported measurement uncertainty:

If the z_U -score of a result is within the tolerance limit and the ζ -score is outside, then the measurement uncertainty is underestimated.

If the z_U -score is outside the tolerance limits and the absolute value of the ζ -score is lower than two, then the requirements of the proficiency test were stronger compared with the reported measurement uncertainty.

Diagrams of uncertainty data

In the first figure for all lab codes the measurement uncertainty (together with the reproducibility standard deviation) is illustrated. The second figure shows the associated ζ -scores.

Measurement uncertainty

10 (40%) out of 25 laboratories with valid values reported measurement uncertainties. In total 231 (37.2%) out of 621 valid values were given with the measurement uncertainty. The following table displays the number of values with measurement uncertainty against the accreditation status.

Accreditation status of the values	Number of values	Number of values with measurement uncertainty
accredited	339	192 (56.6%)
not accredited	66	33 (50%)
not specified	216	6 (2.8%)

We would like to put emphasis on the fact that reporting of measurement uncertainties in our PT scheme is absolutely voluntary. The only objective is to help all participants to reasonably handle measurement uncertainties and their estimation.

The diagrams show that the spread of reported uncertainties in some cases is vast, from unrealistic low values up to very high. A plausibility check using reproducibility standard deviations of the PT round could be helpful here.

If measurement uncertainties are underestimated values assessed as “satisfactory” in the PT ($|z_U| \leq 2$), will have a large ζ -score. $|\zeta| > 2$ means that the “own” requirements (defined in terms of estimated uncertainty) are not fulfilled.

27 (13.9%) of the 194 values reported with uncertainties and having a z_U -score $|z_U| \leq 2.0$ had a ζ -score > 2.0 . This means that the requirements of the PT scheme have been fulfilled, but not the “own” requirements, the uncertainty is underestimated.

Traceable reference values

Traceability of analytical results to national and international references is of increasing importance in all laboratories. This is not easy to realise for chemical analyses and often can only be done by analysing certified reference materials. But availability of these reference materials in the water sector is very limited. Therefore we try to provide reference values for the proficiency test samples, traceable to national and international references.

Since our PT samples without exception are spiked, real water samples, reference values can be calculated from the sum of matrix content and spiked amount of analyte. For both summands traceable values and their uncertainty have to be determined. Thereby we assume that no unrecognised bias resulting from sample preparation and transport is present and that we recognise all uncertainty components.

All spiking of samples was controlled gravimetrically and volumetrically. This procedure allows the preparation of a complete uncertainty budget.


The first step is the specification of the measurand with a formula. This shows the links between the result and all influence quantities for the parameter trimethoprim.

$$c_{\text{lot}} = \frac{m_{\text{ss}} \cdot P \cdot V_{\text{d1}} \cdot V_{\text{d2}} \cdot V_{\text{lot}} \cdot \rho_{\text{lot}}}{V_{\text{ss}} \cdot V_{\text{d1_total}} \cdot V_{\text{d2_total}} \cdot m_{\text{lot_total}}}$$

with:

c_{lot}	concentration of the analyte in the lot resulting from the spike
m_{ss}	mass of substance added for preparation of the stock solution
V_{d1}	volume of stock solution added into the dilution A
V_{d2}	volume of dilution A added into the dilution B
V_{lot}	volume of dilution B added into the lot
V_{ss}	volume of stock solution
$V_{\text{d1_total}}$	total volume of dilution A
$V_{\text{d2_total}}$	total volume of dilution B
$m_{\text{lot_total}}$	total mass of the lot
ρ_{lot}	density of the lot in g/l
P	purity of the used substance

Based on this formula the uncertainty budget can be prepared and all components can be quantified. The following figure shows a typical distribution of the contributions for trimethoprim. The main contributions result from the expanded uncertainty (0,5%), of the purity of the chemical (99,1%) and the pipette steps.



Attention was paid to use a ground water which did not contain any of the analytes. Therefore no matrix content had to be considered and the reference values could be calculated directly from the spikes.


Internet

The report is available on the following webpage: <http://www.aqsbw/pdf/report516.pdf>

sulfadiazine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1205	1,80	0,0100	0,0100	8,31	0,1415	0,1013	17,43	-15,94	21	1	3	19,0
2	0,4518	1,80	0,0519	0,0519	11,49	0,5626	0,3534	24,53	-21,77	21	0	4	19,0
3	0,8133	1,80	0,1008	0,1008	12,39	1,029	0,6231	26,58	-23,39	21	1	1	9,5
sum										63	2	8	15,9

Recovery




slope of the regression: 0,975
 average recovery: 97,5%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	measured data			reference data		
	mean [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]	reference value [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]
1	0,1116	0,0055	4,9	0,1205	0,0022	1,8
2	0,4353	0,0283	6,5	0,4518	0,0082	1,8
3	0,7855	0,0550	7,0	0,8133	0,0147	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,113	0,004	3,958	0,014	12,26	15	1	3	26,67
2	0,454	0,028	6,075	0,085	18,82	15	1	1	13,33
3	0,79	0,04	5,124	0,125	15,88	15	1	1	13,33

sulfadimidine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1223	1,88	0,0189	0,0189	15,48	0,1638	0,0871	33,97	-28,80	24	0	3	13,0
2	0,2935	1,90	0,0284	0,0284	9,66	0,3535	0,2394	20,44	-18,42	23	4	2	26,1
3	0,6604	1,91	0,0575	0,0575	8,71	0,7813	0,5502	18,32	-16,67	22	3	1	17,4
sum										69	7	6	18,8

Recovery




slope of the regression: 0,919
 average recovery: 91,9%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	0,1142	0,0097	8,5	0,1244	0,0023	1,9
2	0,2716	0,0148	5,4	0,2956	0,0056	1,9
3	0,6083	0,0307	5,0	0,6624	0,0126	1,9








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,114	0,005	4,269	0,016	13,66	16	1	2	18,75
2	0,278	0,009	3,269	0,029	10,46	16	1	2	18,75
3	0,602	0,016	2,7	0,052	8,639	16	1	2	18,75

sulfamethoxazole


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1081	1,83	0,0128	0,0128	11,87	0,1356	0,0838	25,40	-22,45	25	0	1	4,0
2	0,3243	1,84	0,0340	0,0340	10,50	0,3965	0,2596	22,26	-19,97	25	0	0	0,0
3	0,7298	1,84	0,0582	0,0582	7,98	0,8514	0,6178	16,68	-15,34	25	2	1	12,0
sum										75	2	2	5,3

Recovery




slope of the regression: 0,955
 average recovery: 95,5%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	measured data			reference data		
	mean [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]	reference value [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]
1	0,1037	0,0064	6,2	0,1088	0,0020	1,8
2	0,3124	0,0170	5,4	0,3250	0,0060	1,8
3	0,6952	0,0291	4,2	0,7304	0,0135	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,103	0,003	3,306	0,012	11,22	18	1	0	5,556
2	0,309	0,011	3,63	0,038	12,32	18	1	0	5,556
3	0,69	0,017	2,456	0,057	8,337	18	1	1	11,11

sulfaethoxy pyridazine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1196	1,76	0,0283	0,0283	23,65	0,1843	0,0681	54,01	-43,06	13	0	1	7,7
2	0,5384	1,81	0,1028	0,1028	19,09	0,7676	0,3489	42,57	-35,20	13	1	0	7,7
3	0,7179	1,82	0,1373	0,1373	19,12	1,023	0,4645	42,57	-35,29	13	0	0	0,0
sum										39	1	1	5,1

Recovery




slope of the regression: 0,1002
 average recovery: 100,2%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	measured values			reference values		
	mean [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]	reference value [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]
1	0,1245	0,0196	15,8	0,1242	0,0022	1,8
2	0,5385	0,0713	13,2	0,5429	0,0099	1,8
3	0,7304	0,0952	13,0	0,7224	0,0131	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,126	0,012	9,299	0,03	23,52	10	0	1	10
2	0,543	0,039	7,114	0,098	18	10	0	0	0
3	0,741	0,071	9,619	0,18	24,33	10	0	0	0

sulfamerazine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1237	1,80	0,0158	0,0158	12,80	0,1578	0,0939	27,55	-24,10	20	0	0	0,0
2	0,3712	1,81	0,0422	0,0422	11,36	0,4612	0,2913	24,25	-21,52	20	2	1	15,0
3	0,6496	1,82	0,0621	0,0621	9,57	0,7808	0,5309	20,20	-18,26	20	2	1	15,0
sum										60	4	2	10,0

Recovery




slope of the regression: 0,943
 average recovery: 94,3%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	measured values			reference values		
	mean [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]	reference value [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]
1	0,1181	0,0089	7,5	0,1252	0,0023	1,8
2	0,3514	0,0236	6,7	0,3727	0,0068	1,8
3	0,6140	0,0347	5,7	0,6511	0,0118	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,121	0,007	5,497	0,019	15,86	13	0	1	7,692
2	0,36	0,016	4,386	0,045	12,65	13	0	1	7,692
3	0,616	0,016	2,624	0,047	7,569	13	1	1	15,38

sulfathiazole


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1248	1,79	0,0108	0,0108	8,63	0,1474	0,1041	18,12	-16,54	19	0	2	10,5
2	0,4678	1,79	0,0512	0,0512	10,95	0,5768	0,3706	23,30	-20,78	19	0	2	10,5
3	0,6550	1,79	0,0491	0,0491	7,50	0,7574	0,5604	15,64	-14,44	19	1	1	10,5
sum										57	1	5	10,5

Recovery




slope of the regression: 0,94
 average recovery: 94,0%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	0,1173	0,0062	5,3	0,1249	0,0022	1,8
2	0,4470	0,0294	6,6	0,4680	0,0084	1,8
3	0,6114	0,0282	4,6	0,6551	0,0117	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,12	0,005	3,831	0,013	11,05	13	0	3	23,08
2	0,445	0,02	4,542	0,058	13,1	13	0	2	15,38
3	0,613	0,017	2,737	0,048	7,895	13	0	1	7,692

sulfadoxine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1290	1,77	0,0194	0,0194	15,03	0,1712	0,0927	32,66	-28,12	16	0	1	6,3
2	0,2903	1,78	0,0266	0,0266	9,16	0,3461	0,2393	19,22	-17,56	16	0	2	12,5
3	0,6774	1,78	0,0821	0,0821	12,12	0,8529	0,5220	25,91	-22,94	16	1	1	12,5
sum										48	1	4	10,4

Recovery




slope of the regression: 0,997
 average recovery: 99,7%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	measured values			reference values		
	mean [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]	reference value [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]
1	0,1291	0,0121	9,4	0,1301	0,0023	1,8
2	0,2926	0,0166	5,7	0,2914	0,0052	1,8
3	0,6712	0,0513	7,6	0,6785	0,0121	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,127	0,008	6,401	0,022	16,98	11	0	1	9,091
2	0,296	0,01	3,248	0,026	8,619	11	0	2	18,18
3	0,68	0,037	5,47	0,099	14,51	11	0	1	9,091

sulfamethoxypridazine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1001	1,86	0,0193	0,0193	19,30	0,1433	0,0646	43,19	-35,50	14	0	2	14,3
2	0,3753	1,87	0,0437	0,0437	11,65	0,4687	0,2925	24,87	-22,06	14	1	1	14,3
3	0,6756	1,88	0,0560	0,0560	8,29	0,7928	0,5679	17,36	-15,94	14	0	1	7,1
sum										42	1	4	11,9

Recovery




slope of the regression: 0,969
 average recovery: 96,9%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	0,0981	0,0129	13,2	0,1012	0,0019	1,9
2	0,3636	0,0292	8,0	0,3764	0,0071	1,9
3	0,6564	0,0374	5,7	0,6767	0,0127	1,9








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,104	0,01	9,295	0,024	23,51	10	0	1	10
2	0,377	0,022	5,817	0,055	14,71	10	0	1	10
3	0,681	0,026	3,776	0,065	9,552	10	0	1	10

sulfachloropyridazine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1141	1,84	0,0266	0,0266	23,30	0,1746	0,0656	53,02	-42,50	15	0	2	13,3
2	0,5134	1,98	0,0741	0,0741	14,43	0,6741	0,3746	31,31	-27,03	15	0	1	6,7
3	0,7701	1,99	0,1616	0,1616	20,99	1,137	0,4748	47,60	-38,34	15	1	1	13,3
sum										45	1	4	11,1

Recovery




slope of the regression: 0,956
 average recovery: 95,6%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	0,1198	0,0172	14,3	0,1254	0,0023	1,8
2	0,5025	0,0478	9,5	0,5247	0,0104	2,0
3	0,7446	0,1043	14,0	0,7814	0,0156	2,0








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,124	0,009	7,184	0,024	19,06	11	0	2	18,18
2	0,521	0,051	9,744	0,135	25,85	11	0	0	0
3	0,777	0,066	8,558	0,176	22,71	11	0	0	0

sulfadimethoxine


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1139	1,78	0,0219	0,0219	19,22	0,1629	0,0737	43,00	-35,36	17	1	2	17,6
2	0,3418	1,81	0,0384	0,0384	11,23	0,4237	0,2690	23,96	-21,29	17	3	0	17,6
3	0,5982	1,82	0,0429	0,0429	7,18	0,6875	0,5153	14,94	-13,86	17	3	1	23,5
sum										51	7	3	19,6

Recovery




slope of the regression: 0,949
 average recovery: 94,9%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values

level	measured values			reference values		
	mean [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]	reference value [$\mu\text{g/l}$]	exp. uncertainty [$\mu\text{g/l}$]	exp. uncertainty [%]
1	0,1115	0,0133	11,9	0,1171	0,0021	1,8
2	0,3247	0,0233	7,2	0,3450	0,0062	1,8
3	0,5720	0,0260	4,6	0,6014	0,0109	1,8








HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,108	0,01	9,315	0,027	24,72	11	0	2	18,18
2	0,301	0,036	11,95	0,095	31,7	11	1	1	18,18
3	0,584	0,036	6,088	0,094	16,15	11	1	0	9,091

trimethoprim


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,1011	1,81	0,0099	0,0099	9,78	0,1220	0,0822	20,61	-18,70	24	0	2	8,3
2	0,3793	1,86	0,0385	0,0385	10,14	0,4605	0,3059	21,42	-19,34	24	0	4	16,7
3	0,6827	1,87	0,0788	0,0788	11,54	0,8509	0,5335	24,64	-21,86	24	2	2	16,7
sum										72	2	8	13,9

Recovery




slope of the regression: 0,969
 average recovery: 96,9%



Relative standard deviation and tolerance limits



The relative standard deviations, calculated with the Q-method, did not reach the limits.






Method specific evaluation




Comparison of means and reference values



level	mean [µg/l]			reference value [µg/l]		
	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	0,1014	0,0050	5,0	0,1048	0,0019	1,8
2	0,3757	0,0196	5,2	0,3829	0,0071	1,9
3	0,6561	0,0402	6,1	0,6863	0,0128	1,9















HPLC-MS/MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	0,1	0,004	3,897	0,013	12,85	17	0	3	17,65
2	0,375	0,015	3,937	0,049	12,99	17	0	3	17,65
3	0,661	0,024	3,614	0,079	11,92	17	1	2	17,65


















PT 5/16 - TW S2		sulfadiazine - 1			
assigned value [$\mu\text{g/l}$]*		0,1205 \pm 0,0022			
upper tolerance limit [$\mu\text{g/l}$]		0,1415			
lower tolerance limit [$\mu\text{g/l}$]		0,1013			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,134			1,3	s
3	0,116			-0,5	s
5	0,1096			-1,1	s
6	0,095	0,024	-2,1	-2,7	q
7	0,108	0,01	-2,4	-1,3	s
9	0,108			-1,3	s
10	0,104			-1,7	s
14	0,15	0,038	1,6	2,8	q
15	0,169	0,042	2,3	4,6	u
16	0,112	0,02	-0,8	-0,9	s
17	0,105			-1,6	s
19	0,109			-1,2	s
20	0,113			-0,8	s
21	0,137	0,055	0,6	1,6	s
23	0,106			-1,5	s
25	0,104	0,021	-1,6	-1,7	s
26	0,111			-1,0	s
31	0,106			-1,5	s
32	0,13			0,9	s
35	0,31			18,1	u
36	0,114	0,013	-1,0	-0,7	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








PT 5/16 - TW S2		sulfadiazine - 2			
assigned value [$\mu\text{g/l}$]*		0,4518 \pm 0,0082			
upper tolerance limit [$\mu\text{g/l}$]		0,5626			
lower tolerance limit [$\mu\text{g/l}$]		0,3534			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,432			-0,4	s
3	0,681			4,1	u
5	0,4032			-1,0	s
6	0,36	0,091	-2,0	-1,9	s
7	0,443	0,087	-0,2	-0,2	s
9	0,458			0,1	s
10	0,384			-1,4	s
14	0,616	0,15	2,2	3,0	u
15	0,587	0,147	1,8	2,4	q
16	0,417	0,06	-1,1	-0,7	s
17	0,438			-0,3	s
19	0,457			0,1	s
20	0,418			-0,7	s
21	0,514	0,206	0,6	1,1	s
23	0,415			-0,7	s
25	0,401	0,08	-1,3	-1,0	s
26	0,405			-1,0	s
31	0,424			-0,6	s
32	0,46			0,1	s
35	1,56			20,0	u
36	0,441	0,02	-1,0	-0,2	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadiazine - 3			
assigned value [$\mu\text{g/l}$]*		0,8133 \pm 0,0147			
upper tolerance limit [$\mu\text{g/l}$]		1,029			
lower tolerance limit [$\mu\text{g/l}$]		0,6231			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,881			0,6	s
3	0,436			-4,0	u
5	0,7651			-0,5	s
6	0,66	0,165	-1,9	-1,6	s
7	0,747	0,133	-1,0	-0,7	s
9	0,825			0,1	s
10	0,692			-1,3	s
14	0,982	0,25	1,3	1,6	s
15	0,915	0,229	0,9	0,9	s
16	0,739	0,12	-1,2	-0,8	s
17	0,724			-0,9	s
19	0,827			0,1	s
20	0,741			-0,8	s
21	0,802	0,321	-0,1	-0,1	s
23	0,825			0,1	s
25	0,719	0,144	-1,3	-1,0	s
26	0,798			-0,2	s
31	0,811			0,0	s
32	0,84			0,2	s
35	2,79			18,3	u
36	0,781	0,083	-0,8	-0,3	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadimidine - 1			
assigned value [$\mu\text{g/l}$]*		0,1223 \pm 0,0023			
upper tolerance limit [$\mu\text{g/l}$]		0,1638			
lower tolerance limit [$\mu\text{g/l}$]		0,08707			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,128			0,3	s
3	0,325			9,8	u
5	0,1061			-0,9	s
6	0,12	0,018	-0,3	-0,1	s
9	0,122			0,0	s
10	0,092			-1,7	s
14	0,177	0,044	2,5	2,6	q
15	0,189	0,028	4,7	3,2	u
16	0,115	0,02	-0,7	-0,4	s
17	0,107			-0,9	s
18	0,119	0,03	-0,2	-0,2	s
19	0,127			0,2	s
20	0,117			-0,3	s
21	0,107	0,043	-0,7	-0,9	s
22	0,103	0,008	-4,6	-1,1	s
23	0,111			-0,6	s
25	0,104	0,021	-1,8	-1,0	s
26	0,127			0,2	s
29	0,106			-0,9	s
31	0,087			-2,0	s
32	0,12			-0,1	s
35	0,116			-0,4	s
36	0,117	0,01	-1,0	-0,3	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadimidine - 2			
assigned value [$\mu\text{g/l}$]*		0,2935 \pm 0,0056			
upper tolerance limit [$\mu\text{g/l}$]		0,3535			
lower tolerance limit [$\mu\text{g/l}$]		0,2394			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,279			-0,5	s
3	0,124			-6,3	u
5	0,2639			-1,1	s
6	0,27	0,04	-1,2	-0,9	s
9	0,29			-0,1	s
10	0,235			-2,2	q
14	0,366	0,092	1,6	2,4	q
15	0,37	0,056	2,7	2,6	q
16	0,27	0,06	-0,8	-0,9	s
17	0,276			-0,6	s
18	0,285	0,071	-0,2	-0,3	s
19	0,292			-0,1	s
20	0,276			-0,6	s
21	0,259	0,104	-0,7	-1,3	s
22	0,244	0,02	-4,8	-1,8	s
23	0,284			-0,4	s
25	0,249	0,05	-1,8	-1,6	s
26	0,305			0,4	s
29	0,225			-2,5	q
31	0,21			-3,1	u
32	0,28			-0,5	s
35	0,279			-0,5	s
36	0,275	0,032	-1,1	-0,7	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadimidine - 3			
assigned value [$\mu\text{g/l}$]*		0,6604 \pm 0,0126			
upper tolerance limit [$\mu\text{g/l}$]		0,7813			
lower tolerance limit [$\mu\text{g/l}$]		0,5502			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,658			0,0	s
3	0,64			-0,4	s
5	0,6233			-0,7	s
6	0,62	0,09	-0,9	-0,7	s
9	0,64			-0,4	s
10	0,54			-2,2	q
14	0,834	0,21	1,7	2,9	q
15	0,768	0,115	1,9	1,8	s
16	0,644	0,1	-0,3	-0,3	s
17	0,55			-2,0	s
18	0,54	0,135	-1,8	-2,2	q
19	0,654			-0,1	s
20	0,629			-0,6	s
21	0,565	0,226	-0,8	-1,7	s
22	0,55	0,044	-4,8	-2,0	s
23	0,672			0,2	s
25	0,566	0,113	-1,7	-1,7	s
26	0,672			0,2	s
29	0,568			-1,7	s
31	0,42			-4,4	u
32	0,66			0,0	s
35	0,557			-1,9	s
36	0,62	0,066	-1,2	-0,7	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamethoxazole - 1			
assigned value [$\mu\text{g/l}$]*		0,1081 \pm 0,002			
upper tolerance limit [$\mu\text{g/l}$]		0,1356			
lower tolerance limit [$\mu\text{g/l}$]		0,08384			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,107			-0,1	s
3	0,632			38,2	u
5	0,0929			-1,3	s
6	0,11	0,012	0,3	0,1	s
7	0,098	0,019	-1,1	-0,8	s
9	0,116			0,6	s
10	0,095			-1,1	s
14	0,127	0,032	1,2	1,4	s
15	0,106	0,016	-0,3	-0,2	s
16	0,101	0,02	-0,7	-0,6	s
17	0,096			-1,0	s
18	0,088	0,022	-1,8	-1,7	s
19	0,099			-0,8	s
20	0,108			0,0	s
21	0,105	0,042	-0,1	-0,3	s
22	0,093	0,003	-8,4	-1,2	s
23	0,12			0,9	s
25	0,092	0,018	-1,7	-1,3	s
26	0,1			-0,7	s
29	0,0889			-1,6	s
31	0,116			0,6	s
32	0,12			0,9	s
33	0,11			0,1	s
35	0,107			-0,1	s
36	0,097	0,009	-2,4	-0,9	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamethoxazole - 2			
assigned value [$\mu\text{g/l}$]*		0,3243 \pm 0,006			
upper tolerance limit [$\mu\text{g/l}$]		0,3965			
lower tolerance limit [$\mu\text{g/l}$]		0,2596			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,328			0,1	s
3	0,332			0,2	s
5	0,2861			-1,2	s
6	0,3	0,034	-1,4	-0,8	s
7	0,308	0,047	-0,7	-0,5	s
9	0,342			0,5	s
10	0,299			-0,8	s
14	0,388	0,097	1,3	1,8	s
15	0,348	0,052	0,9	0,7	s
16	0,311	0,04	-0,7	-0,4	s
17	0,27			-1,7	s
18	0,264	0,066	-1,8	-1,9	s
19	0,318			-0,2	s
20	0,326			0,0	s
21	0,292	0,117	-0,6	-1,0	s
22	0,274	0,008	-10,1	-1,6	s
23	0,389			1,8	s
25	0,287	0,057	-1,3	-1,2	s
26	0,309			-0,5	s
29	0,312			-0,4	s
31	0,318			-0,2	s
32	0,33			0,2	s
33	0,35			0,7	s
35	0,281			-1,3	s
36	0,298	0,025	-2,0	-0,8	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








PT 5/16 - TW S2		sulfamethoxazole - 3			
assigned value [$\mu\text{g/l}$]*		0,7298 \pm 0,0135			
upper tolerance limit [$\mu\text{g/l}$]		0,8514			
lower tolerance limit [$\mu\text{g/l}$]		0,6178			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,7			-0,5	s
3	0,108			-11,1	u
5	0,6523			-1,4	s
6	0,69	0,078	-1,0	-0,7	s
7	0,73	0,175	0,0	0,0	s
9	0,757			0,4	s
10	0,684			-0,8	s
14	0,978	0,24	2,1	4,1	u
15	0,756	0,113	0,5	0,4	s
16	0,689	0,12	-0,7	-0,7	s
17	0,68			-0,9	s
18	0,586	0,147	-1,9	-2,6	q
19	0,71			-0,4	s
20	0,744			0,2	s
21	0,671	0,268	-0,4	-1,0	s
22	0,624	0,019	-9,1	-1,9	s
23	0,749			0,3	s
25	0,646	0,129	-1,3	-1,5	s
26	0,657			-1,3	s
29	0,702			-0,5	s
31	0,728			0,0	s
32	0,71			-0,4	s
33	0,76			0,5	s
35	0,667			-1,1	s
36	0,675	0,044	-2,4	-1,0	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfaethoxypyridazine - 1			
assigned value [$\mu\text{g/l}$]*		0,1196 \pm 0,0022			
upper tolerance limit [$\mu\text{g/l}$]		0,1843			
lower tolerance limit [$\mu\text{g/l}$]		0,06813			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,518			12,3	u
5	0,1242			0,1	s
9	0,122			0,1	s
10	0,084			-1,4	s
14	0,158	0,04	1,9	1,2	s
15	0,146	0,022	2,4	0,8	s
16	0,143	0,02	2,3	0,7	s
17	0,104			-0,6	s
19	0,127			0,2	s
25	0,116	0,023	-0,3	-0,1	s
26	0,132			0,4	s
29	0,103			-0,6	s
36	0,135	0,009	3,3	0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfaethoxypyridazine - 2			
assigned value [$\mu\text{g/l}$]*		0,5384 \pm 0,0099			
upper tolerance limit [$\mu\text{g/l}$]		0,7676			
lower tolerance limit [$\mu\text{g/l}$]		0,3489			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,116			-4,5	u
5	0,57			0,3	s
9	0,531			-0,1	s
10	0,424			-1,2	s
14	0,634	0,16	1,2	0,8	s
15	0,628	0,094	1,9	0,8	s
16	0,626	0,18	1,0	0,8	s
17	0,392			-1,5	s
19	0,536			0,0	s
25	0,531	0,106	-0,1	-0,1	s
26	0,579			0,4	s
29	0,448			-1,0	s
36	0,603	0,053	2,4	0,6	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfaethoxypyridazine - 3			
assigned value [$\mu\text{g/l}$]*		0,7179 \pm 0,0131			
upper tolerance limit [$\mu\text{g/l}$]		1,023			
lower tolerance limit [$\mu\text{g/l}$]		0,4645			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,62			-0,8	s
5	0,7647			0,3	s
9	0,697			-0,2	s
10	0,568			-1,2	s
14	0,856	0,21	1,3	0,9	s
15	0,856	0,128	2,1	0,9	s
16	0,895	0,28	1,3	1,2	s
17	0,582			-1,1	s
19	0,698			-0,2	s
25	0,714	0,143	-0,1	0,0	s
26	0,8			0,5	s
29	0,649			-0,5	s
36	0,795	0,08	1,9	0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamerazine - 1			
assigned value [$\mu\text{g/l}$]*		0,1237 \pm 0,0023			
upper tolerance limit [$\mu\text{g/l}$]		0,1578			
lower tolerance limit [$\mu\text{g/l}$]		0,0939			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,129			0,3	s
3	0,128			0,3	s
5	0,1112			-0,8	s
6	0,13	0,025	0,5	0,4	s
9	0,124			0,0	s
10	0,092			-2,1	q
14	0,154	0,038	1,6	1,8	s
15	0,15	0,022	2,4	1,5	s
16	0,115	0,02	-0,9	-0,6	s
17	0,104			-1,3	s
19	0,122			-0,1	s
20	0,115			-0,6	s
21	0,126	0,05	0,1	0,1	s
23	0,103			-1,4	s
25	0,105	0,021	-1,8	-1,3	s
26	0,114			-0,7	s
29	0,105			-1,3	s
31	0,114			-0,7	s
35	0,123			0,0	s
36	0,116	0,012	-1,3	-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamerazine - 2			
assigned value [$\mu\text{g/l}$]*		0,3712 \pm 0,0068			
upper tolerance limit [$\mu\text{g/l}$]		0,4612			
lower tolerance limit [$\mu\text{g/l}$]		0,2913			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,372			0,0	s
3	0,58			4,6	u
5	0,3358			-0,9	s
6	0,32	0,062	-1,6	-1,3	s
9	0,36			-0,3	s
10	0,279			-2,3	q
14	0,442	0,11	1,3	1,6	s
15	0,381	0,057	0,3	0,2	s
16	0,337	0,06	-1,1	-0,9	s
17	0,382			0,2	s
19	0,368			-0,1	s
20	0,341			-0,8	s
21	0,363	0,145	-0,1	-0,2	s
23	0,334			-0,9	s
25	0,318	0,064	-1,7	-1,3	s
26	0,364			-0,2	s
29	0,275			-2,4	q
31	0,357			-0,4	s
35	0,403			0,7	s
36	0,35	0,04	-1,0	-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamerazine - 3			
assigned value [$\mu\text{g/l}$]*		0,6496 \pm 0,0118			
upper tolerance limit [$\mu\text{g/l}$]		0,7808			
lower tolerance limit [$\mu\text{g/l}$]		0,5309			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,653			0,1	s
3	0,319			-5,6	u
5	0,5799			-1,2	s
6	0,61	0,12	-0,7	-0,7	s
9	0,625			-0,4	s
10	0,483			-2,8	q
14	0,85	0,21	1,9	3,1	u
15	0,674	0,101	0,5	0,4	s
16	0,609	0,1	-0,8	-0,7	s
17	0,613			-0,6	s
19	0,654			0,1	s
20	0,591			-1,0	s
21	0,597	0,259	-0,4	-0,9	s
23	0,571			-1,3	s
25	0,564	0,113	-1,5	-1,4	s
26	0,618			-0,5	s
29	0,556			-1,6	s
31	0,681			0,5	s
35	0,663			0,2	s
36	0,629	0,031	-1,2	-0,3	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








PT 5/16 - TW S2		sulfathiazole - 1			
assigned value [$\mu\text{g/l}$]*		0,1248 \pm 0,0022			
upper tolerance limit [$\mu\text{g/l}$]		0,1474			
lower tolerance limit [$\mu\text{g/l}$]		0,1041			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,127			0,2	s
3	0,121			-0,4	s
5	0,1089			-1,5	s
6	0,13	0,025	0,4	0,5	s
9	0,124			-0,1	s
10	0,105			-1,9	s
14	0,165	0,041	2,0	3,6	u
15	0,158	0,039	1,7	2,9	q
16	0,114	0,02	-1,1	-1,0	s
17	0,12			-0,5	s
18	0,111	0,028	-1,0	-1,3	s
19	0,124			-0,1	s
23	0,113			-1,1	s
25	0,107	0,022	-1,6	-1,7	s
26	0,114			-1,0	s
29	0,115			-0,9	s
31	0,112			-1,2	s
32	0,12			-0,5	s
36	0,12	0,008	-1,1	-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








PT 5/16 - TW S2		sulfathiazole - 2			
assigned value [$\mu\text{g/l}$]*		0,4678 \pm 0,0084			
upper tolerance limit [$\mu\text{g/l}$]		0,5768			
lower tolerance limit [$\mu\text{g/l}$]		0,3706			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,466			0,0	s
3	0,643			3,2	u
5	0,4307			-0,8	s
6	0,52	0,102	1,0	1,0	s
9	0,464			-0,1	s
10	0,387			-1,7	s
14	0,636	0,16	2,1	3,1	u
15	0,437	0,109	-0,6	-0,6	s
16	0,441	0,06	-0,9	-0,6	s
17	0,492			0,4	s
18	0,402	0,101	-1,3	-1,4	s
19	0,379			-1,8	s
23	0,42			-1,0	s
25	0,413	0,083	-1,3	-1,1	s
26	0,437			-0,6	s
29	0,443			-0,5	s
31	0,483			0,3	s
32	0,45			-0,4	s
36	0,458	0,025	-0,7	-0,2	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfathiazole - 3			
assigned value [$\mu\text{g/l}$]*		0,655 \pm 0,0117			
upper tolerance limit [$\mu\text{g/l}$]		0,7574			
lower tolerance limit [$\mu\text{g/l}$]		0,5604			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,601			-1,1	s
3	0,457			-4,2	u
5	0,5863			-1,5	s
6	0,64	0,127	-0,2	-0,3	s
9	0,652			-0,1	s
10	0,565			-1,9	s
14	0,808	0,2	1,5	3,0	u
15	0,651	0,163	0,0	-0,1	s
16	0,628	0,12	-0,4	-0,6	s
17	0,62			-0,7	s
18	0,59	0,148	-0,9	-1,4	s
19	0,572			-1,8	s
23	0,632			-0,5	s
25	0,577	0,115	-1,3	-1,6	s
26	0,623			-0,7	s
29	0,57			-1,8	s
31	0,734			1,5	s
32	0,6			-1,2	s
36	0,644	0,064	-0,3	-0,2	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadoxine - 1			
assigned value [$\mu\text{g/l}$]*		0,129 \pm 0,0023			
upper tolerance limit [$\mu\text{g/l}$]		0,1712			
lower tolerance limit [$\mu\text{g/l}$]		0,09274			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,133			0,2	s
3	0,719			28,0	u
5	0,1191			-0,5	s
9	0,134			0,2	s
10	0,104			-1,4	s
14	0,157	0,039	1,4	1,3	s
15	0,141	0,035	0,7	0,6	s
16	0,118	0,02	-1,1	-0,6	s
17	0,129			0,0	s
19	0,119			-0,6	s
23	0,148			0,9	s
25	0,11	0,022	-1,7	-1,0	s
26	0,122			-0,4	s
29	0,134			0,2	s
32	0,14			0,5	s
36	0,128	0,016	-0,1	-0,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadoxine - 2			
assigned value [$\mu\text{g/l}$]*		0,2903 \pm 0,0052			
upper tolerance limit [$\mu\text{g/l}$]		0,3461			
lower tolerance limit [$\mu\text{g/l}$]		0,2393			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,282			-0,3	s
3	0,36			2,5	q
5	0,2714			-0,7	s
9	0,294			0,1	s
10	0,247			-1,7	s
14	0,36	0,09	1,5	2,5	q
15	0,313	0,078	0,6	0,8	s
16	0,312	0,06	0,7	0,8	s
17	0,302			0,4	s
19	0,308			0,6	s
23	0,274			-0,6	s
25	0,253	0,051	-1,5	-1,5	s
26	0,282			-0,3	s
29	0,29			0,0	s
32	0,3			0,3	s
36	0,283	0,029	-0,5	-0,3	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadoxine - 3			
assigned value [$\mu\text{g/l}$]*		0,6774 \pm 0,0121			
upper tolerance limit [$\mu\text{g/l}$]		0,8529			
lower tolerance limit [$\mu\text{g/l}$]		0,522			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,673			-0,1	s
3	0,13			-7,0	u
5	0,6313			-0,6	s
9	0,684			0,1	s
10	0,579			-1,3	s
14	0,923	0,23	2,1	2,8	q
15	0,734	0,183	0,6	0,6	s
16	0,617	0,18	-0,7	-0,8	s
17	0,738			0,7	s
19	0,649			-0,4	s
23	0,675			0,0	s
25	0,596	0,119	-1,4	-1,0	s
26	0,639			-0,5	s
29	0,641			-0,5	s
32	0,74			0,7	s
36	0,683	0,026	0,4	0,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamethoxypyridazine - 1			
assigned value [$\mu\text{g/l}$]*		0,1001 \pm 0,0019			
upper tolerance limit [$\mu\text{g/l}$]		0,1433			
lower tolerance limit [$\mu\text{g/l}$]		0,06456			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,351			11,6	u
5	0,0891			-0,6	s
9	0,108			0,4	s
10	0,078			-1,2	s
14	0,145	0,036	2,5	2,1	q
15	0,134	0,02	3,4	1,6	s
16	0,096	0,02	-0,4	-0,2	s
17	0,093			-0,4	s
19	0,097			-0,2	s
23	0,093			-0,4	s
25	0,0829	0,017	-2,1	-1,0	s
26	0,105			0,2	s
29	0,0813			-1,1	s
36	0,098	0,007	-0,6	-0,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamethoxypyridazine - 2			
assigned value [$\mu\text{g/l}$]*		0,3753 \pm 0,0071			
upper tolerance limit [$\mu\text{g/l}$]		0,4687			
lower tolerance limit [$\mu\text{g/l}$]		0,2925			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,104			-6,6	u
5	0,3456			-0,7	s
9	0,395			0,4	s
10	0,352			-0,6	s
14	0,48	0,12	1,7	2,2	q
15	0,419	0,063	1,4	0,9	s
16	0,353	0,06	-0,7	-0,5	s
17	0,339			-0,9	s
19	0,368			-0,2	s
23	0,342			-0,8	s
25	0,326	0,065	-1,5	-1,2	s
26	0,381			0,1	s
29	0,315			-1,5	s
36	0,362	0,044	-0,6	-0,3	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfamethoxypyridazine - 3			
assigned value [$\mu\text{g/l}$]*		0,6756 \pm 0,0127			
upper tolerance limit [$\mu\text{g/l}$]		0,7928			
lower tolerance limit [$\mu\text{g/l}$]		0,5679			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,639			-0,7	s
5	0,6227			-1,0	s
9	0,693			0,3	s
10	0,623			-1,0	s
14	0,861	0,22	1,7	3,2	u
15	0,764	0,115	1,5	1,5	s
16	0,662	0,12	-0,2	-0,3	s
17	0,667			-0,2	s
19	0,67			-0,1	s
23	0,655			-0,4	s
25	0,587	0,117	-1,5	-1,6	s
26	0,688			0,2	s
29	0,594			-1,5	s
36	0,644	0,074	-0,8	-0,6	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfachloropyridazine - 1			
assigned value [$\mu\text{g/l}$]*		0,1141 \pm 0,0023			
upper tolerance limit [$\mu\text{g/l}$]		0,1746			
lower tolerance limit [$\mu\text{g/l}$]		0,0656			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,709			19,7	u
5	0,0985			-0,6	s
6	0,15	0,019	3,8	1,2	s
9	0,18			2,2	q
10	0,098			-0,7	s
14	0,146	0,036	1,8	1,1	s
15	0,147	0,015	4,3	1,1	s
16	0,113	0,02	-0,1	0,0	s
17	0,101			-0,5	s
19	0,117			0,1	s
23	0,126			0,4	s
25	0,0994	0,02	-1,5	-0,6	s
26	0,109			-0,2	s
29	0,1			-0,6	s
36	0,113	0,015	-0,1	0,0	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfachloropyridazine - 2			
assigned value [$\mu\text{g/l}$]*		0,5134 \pm 0,0104			
upper tolerance limit [$\mu\text{g/l}$]		0,6741			
lower tolerance limit [$\mu\text{g/l}$]		0,3746			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,487			-0,4	s
5	0,4715			-0,6	s
6	0,39	0,049	-4,9	-1,8	s
9	0,763			3,1	u
10	0,451			-0,9	s
14	0,636	0,16	1,5	1,5	s
15	0,592	0,059	2,6	1,0	s
16	0,523	0,1	0,2	0,1	s
17	0,466			-0,7	s
19	0,524			0,1	s
23	0,552			0,5	s
25	0,45	0,09	-1,4	-0,9	s
26	0,459			-0,8	s
29	0,476			-0,5	s
36	0,506	0,07	-0,2	-0,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfachloropyridazine - 3			
assigned value [$\mu\text{g/l}$]*		0,7701 \pm 0,0156			
upper tolerance limit [$\mu\text{g/l}$]		1,137			
lower tolerance limit [$\mu\text{g/l}$]		0,4748			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
3	0,118			-4,4	u
5	0,6369			-0,9	s
6	0,54	0,067	-6,7	-1,6	s
9	1,11			1,9	s
10	0,638			-0,9	s
14	1,039	0,26	2,1	1,5	s
15	0,885	0,089	2,5	0,6	s
16	0,742	0,1	-0,6	-0,2	s
17	0,736			-0,2	s
19	0,777			0,0	s
23	0,789			0,1	s
25	0,661	0,132	-1,6	-0,7	s
26	0,71			-0,4	s
29	0,677			-0,6	s
36	0,759	0,066	-0,3	-0,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadimethoxine - 1			
assigned value [$\mu\text{g/l}$]*		0,1139 \pm 0,0021			
upper tolerance limit [$\mu\text{g/l}$]		0,1629			
lower tolerance limit [$\mu\text{g/l}$]		0,07365			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,117			0,1	s
3	0,583			19,1	u
5	0,1032			-0,5	s
6	0,069	0,012	-7,4	-2,2	q
9	0,122			0,3	s
10	0,102			-0,6	s
14	0,136	0,034	1,3	0,9	s
15	0,131	0,02	1,7	0,7	s
16	0,11	0,02	-0,4	-0,2	s
17	0,098			-0,8	s
19	0,344			9,4	u
23	0,103			-0,5	s
25	0,0977	0,02	-1,7	-0,8	s
26	0,108			-0,3	s
29	0,151			1,5	s
31	0,113			0,0	s
36	0,109	0,01	-1,0	-0,2	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadimethoxine - 2			
assigned value [$\mu\text{g/l}$]*		0,3418 \pm 0,0062			
upper tolerance limit [$\mu\text{g/l}$]		0,4237			
lower tolerance limit [$\mu\text{g/l}$]		0,269			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,334			-0,2	s
3	0,33			-0,3	s
5	0,3076			-0,9	s
6	0,19	0,032	-9,3	-4,2	u
9	0,354			0,3	s
10	0,335			-0,2	s
14	0,396	0,099	1,1	1,3	s
15	0,379	0,057	1,3	0,9	s
16	0,334	0,06	-0,3	-0,2	s
17	0,261			-2,2	q
19	0,118			-6,2	u
23	0,327			-0,4	s
25	0,29	0,058	-1,8	-1,4	s
26	0,313			-0,8	s
29	0,317			-0,7	s
31	0,314			-0,8	s
36	0,325	0,02	-1,6	-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		sulfadimethoxine - 3			
assigned value [$\mu\text{g/l}$]*		0,5982 \pm 0,0109			
upper tolerance limit [$\mu\text{g/l}$]		0,6875			
lower tolerance limit [$\mu\text{g/l}$]		0,5153			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,573			-0,6	s
3	0,111			-11,8	u
5	0,5562			-1,0	s
6	0,29	0,048	-12,5	-7,4	u
9	0,618			0,4	s
10	0,576			-0,5	s
14	0,784	0,2	1,9	4,2	u
15	0,664	0,1	1,3	1,5	s
16	0,562	0,1	-0,7	-0,9	s
17	0,572			-0,6	s
19	0,6			0,0	s
23	0,57			-0,7	s
25	0,51	0,102	-1,7	-2,1	q
26	0,558			-1,0	s
29	0,522			-1,8	s
31	0,567			-0,8	s
36	0,588	0,022	-0,8	-0,2	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








PT 5/16 - TW S2		trimethoprim - 1			
assigned value [$\mu\text{g/l}$]*		0,1011 \pm 0,0019			
upper tolerance limit [$\mu\text{g/l}$]		0,122			
lower tolerance limit [$\mu\text{g/l}$]		0,08222			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,109			0,8	s
3	0,11			0,9	s
5	0,1052			0,4	s
6	0,081	0,014	-2,9	-2,1	q
7	0,098	0,044	-0,1	-0,3	s
9	0,114			1,2	s
10	0,101			0,0	s
14	0,181	0,045	3,5	7,7	u
15	0,211	0,032	6,9	10,5	u
16	0,103	0,04	0,1	0,2	s
17	0,083			-1,9	s
18	0,101	0,025	0,0	0,0	s
19	0,103			0,2	s
20	0,104			0,3	s
21	0,115	0,046	0,6	1,3	s
23	0,093			-0,9	s
25	0,0874	0,018	-1,6	-1,5	s
26	0,103			0,2	s
29	0,102			0,1	s
31	0,102			0,1	s
32	0,1			-0,1	s
33	0,11			0,9	s
35	0,101			0,0	s
36	0,096	0,005	-1,9	-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory









PT 5/16 - TW S2		trimethoprim - 2			
assigned value [$\mu\text{g/l}$]*		0,3793 \pm 0,0071			
upper tolerance limit [$\mu\text{g/l}$]		0,4605			
lower tolerance limit [$\mu\text{g/l}$]		0,3059			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,407			0,7	s
3	0,684			7,5	u
5	0,4073			0,7	s
6	0,33	0,059	-1,7	-1,3	s
7	0,368	0,068	-0,3	-0,3	s
9	0,402			0,6	s
10	0,378			0,0	s
14	0,626	0,16	3,1	6,1	u
15	0,659	0,099	5,6	6,9	u
16	0,367	0,12	-0,2	-0,3	s
17	0,378			0,0	s
18	0,333	0,083	-1,1	-1,3	s
19	0,362			-0,5	s
20	0,382			0,1	s
21	0,416	0,166	0,4	0,9	s
23	0,354			-0,7	s
25	0,337	0,067	-1,2	-1,2	s
26	0,367			-0,3	s
29	0,361			-0,5	s
31	0,386			0,2	s
32	0,39			0,3	s
33	0,48			2,5	q
35	0,368			-0,3	s
36	0,364	0,013	-2,1	-0,4	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory





PT 5/16 - TW S2		trimethoprim - 3			
assigned value [$\mu\text{g/l}$]*		0,6827 \pm 0,0128			
upper tolerance limit [$\mu\text{g/l}$]		0,8509			
lower tolerance limit [$\mu\text{g/l}$]		0,5335			
lab code	result [$\mu\text{g/l}$]	\pm	ζ -score	z_U -score	assessm.**
2	0,635			-0,6	s
3	0,368			-4,2	u
5	0,7383			0,7	s
6	0,49	0,087	-4,4	-2,6	q
7	0,705	0,194	0,2	0,3	s
9	0,721			0,5	s
10	0,669			-0,2	s
14	1	0,25	2,5	3,8	u
15	1,064	0,16	4,8	4,5	u
16	0,664	0,22	-0,2	-0,3	s
17	0,533			-2,0	s
18	0,684	0,171	0,0	0,0	s
19	0,649			-0,5	s
20	0,699			0,2	s
21	0,709	0,284	0,2	0,3	s
23	0,621			-0,8	s
25	0,603	0,121	-1,3	-1,1	s
26	0,671			-0,2	s
29	0,551			-1,8	s
31	0,716			0,4	s
32	0,65			-0,4	s
33	0,72			0,4	s
35	0,706			0,3	s
36	0,646	0,034	-2,0	-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor $k=2$ corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory

